



Example of how to set up a defect Library

Defect Type	Defect Description	Illustration of Vial Defect
Critical	Broken – Fragmented into separate parts and causes leakage	
Critical	Cracked – Fissured without separation of parts. (Any surface flaw that compromises sealing and overall product integrity.) (NOTE: During AQL, questionable surface flaws may be submitted to QC for CCIT)	

Repeat this process with all Critical Major and minor defects